



NOTES:

1. LEAK TEST: 80 PSI, NO BUBBLES 30 SECONDS MINIMUM.
2. HYPOT 630 VDC 500 MEGOHMS .01 SEC MINIMUM, WIRE TO WIRE & HOUSING.
3. FOR QUALIFICATION ONLY:
 - a. HELIUM LEAK TEST: 10^{-8} cc/s
 - b. HYPOT 1000 VAC WIRE TO WIRE & HOUSING .5 mA MAX LEAKAGE.
4. THIS IS A UL/CSA IMPACT COMPONENT. AFTER APPROVAL BY BECKMAN ENGINEERING NO CHANGES ARE TO BE MADE IN MATERIAL, PROCEDURES, OR TEST REQUIREMENTS.
5. ALL TESTS ARE DONE AT ROOM TEMPERATURE.
6. THESE SPECIFICATIONS SUPERSEDE PAVE CATALOG.

3	2.47 FT	EE22 UL1180 BROWN	WIRE 22EE UL1180 .016 7/#30
2	A/R	PAVE-Seal 80	EPOXY BLACK
1	1	0120-SS	HOUSING PT2-PS
ITEM	QTY	PART#	DESCRIPTION

ALL DIMENSIONS AND TOLERANCES APPLY TO FINISHED PART IN INCHES
 ALLOWABLE TOLERANCES UNLESS SPECIFIED OTHERWISE: NONE +/- .5
 .X DECIMAL +/- .1 .XX DECIMAL +/- .02 .XXX DECIMAL +/- .005
 ANGLES +/- 1 DEG SURFACE FINISH 128 microinch RMS

PAVE technology

2751 Thunderhawk Court
 Dayton, Ohio 45414-3445
 U.S.A.
 (937)890-1100
 fax (937)890-5165
 www.pavetechnologyco.com

DESCRIPTION PAVE-Seal® Cable Harnesses

PT2-SS-80-2-TEE22-20-8

PART NUMBER	PROJECTION
0900	