



- NOTES:
1. HE LEAK TEST @1 ATM. 10^{-8} CC/S.
 2. HYPOT 630 VDC 500 MW 0.01 SEC MINIMUM, WIRE TO WIRE & HOUSING.
 3. ALL TESTS ARE PERFORMED AT ROOM TEMPERATURE.

PART #	X	QTY
1581	48	12@98"
1581-1	96	12@146"

3	A/R	PAVE-Seal 150	EPOXY BLACK
2	SEE TABLE	KT20	TYPE K #20 TEFLON
1	1	NW40-SS	HOUSING NW40-SS
ITEM	QTY	PART#	DESCRIPTION

ALL DIMENSIONS AND TOLERANCES APPLY TO FINISHED PART IN INCHES
 ALLOWABLE TOLERANCES UNLESS SPECIFIED OTHERWISE: NONE +/-0.5
 .X DECIMAL +/- 0.1 .XX DECIMAL +/- 0.02 .XXX DECIMAL +/-0.005
 ANGLES +/- 1 DEG SURFACE FINISH 128 microinch RMS

2751 Thunderhawk Court
 Dayton, Ohio 45414-3445
 U.S.A.
 (937)890-1100
 Fax (937)890-5165
 www.pavetechnologyco.com

PAVE Technology Co.

DESCRIPTION PAVE-Seal@ Cable Harnesses
NW40-SS-150-12P-KT20-X-48

PART NUMBER
1581

PROJECTION