



NOTES:

1. HE LEAK TEST @1 ATM. $<10^{-6}$ CC/S.
2. HYPOT 630 VDC 500 M Ω .01 SEC MINIMUM, WIRE TO WIRE & HOUSING.
3. ALL TESTS ARE PERFORMED AT ROOM TEMPERATURE.
4. FINISH APPLIES TO SPOTFACE FOR O-RING SEAL ONLY.

6	A/R	SUPER-O-LUBE	GREASE O-RING SILICONE BASED
5	A/R	PAVESEAL 150	EPOXY BLACK
4	1	0183	NUT VS15-SS
3	1	-215 VITON 75A	O-RING -215 VITON 75A
2	1	0161	HOUSING VS15L-SS
1	9	JT20	CABLE T/C 20 TYPE J TEFLON
ITEM	QTY	PART NUMBER	DESCRIPTION

ALL DIMENSIONS AND TOLERANCES APPLY TO FINISHED PART IN INCHES
 ALLOWABLE TOLERANCES UNLESS SPECIFIED OTHERWISE: NONE +/-0.5
 .X DECIMAL +/- 0.1 .XX DECIMAL +/- 0.02 .XXX DECIMAL +/-0.005
 ANGLES +/- 1 DEG SURFACE FINISH 128 microinch RMS

2751 Thunderhawk Court
 Dayton, Ohio 45414-3445
 U.S.A.
 (937)890-1100
 fax (937)890-5165
 www.pavetechnologyco.com

PAVE technology co.

DESCRIPTION
VS15L-SS-150-9P-JT20-48-48

PART NUMBER
1664

REVISION LEVEL
B

PROJECTION