



NOTES:

1. HE LEAK TEST @1 ATM. 10×10^{-7} CC/S OR LESS.
2. HYPOT 630 VDC 500 MΩ 0.01 SEC MINIMUM, PIN TO PIN & HOUSING.
3. ALL TESTS ARE PERFORMED AT ROOM TEMPERATURE.
4. ALL PARTS MUST PASS ALL TESTS.
5. DIMENSIONS ARE INCHES [millimeters].

5	3	M23053/8-005-C	TUBING HEAT SHRINK .010 KYNAR ø.187
4	A/R	PAVE-Seal 150	EPOXY BLACK
3	3	M39029/4-111	MALE #16 CONTACT
2	1	SC3112E12-3P	RCPT 3112 SHELL 12 3#16
1	1	NW-25FL	HOUSING VACUUM FLANGE
ITEM	QTY	PART#	DESCRIPTION

ALL DIMENSIONS AND TOLERANCES APPLY TO FINISHED PART IN INCHES
 ALLOWABLE TOLERANCES UNLESS SPECIFIED OTHERWISE: NONE +/-0.5
 .X DECIMAL +/- 0.1 .XX DECIMAL +/- 0.02 .XXX DECIMAL +/-0.005
 ANGLES +/- 1 DEG SURFACE FINISH 128 microinch RMS

2751 Thunderhawk Court
Dayton, Ohio 45414-3445
U.S.A.
(937)890-1100
fax (937)890-5165
www.pavetechnologyco.com

PAVE technology co.

DESCRIPTION PAVE-Mate II Connectors
 NW25-SS-150-3-16-3112E12-3P-P

PART NUMBER **1694**

REVISION LEVEL **E**

PROJECTION