



NOTES:

1. HE LEAK TEST @1 ATM. 10^{-6} CC/S.
2. HYPOT 630 VDC 500 MW 0.01 SEC MINIMUM, WIRE TO WIRE & HOUSING.
3. ALL TESTS ARE PERFORMED AT ROOM TEMPERATURE.

3	A/R	PAVE-Seal 150	EPOXY BLACK
2	12@98"	BXT20	CABLE T/C TYPE BX 20AWG TEFLON
1	1	NW40-SS	HOUSING NW40-SS
ITEM	QTY	PART#	DESCRIPTION

ALL DIMENSIONS AND TOLERANCES APPLY TO FINISHED PART IN INCHES
 ALLOWABLE TOLERANCES UNLESS SPECIFIED OTHERWISE: NONE +/-0.5
 .X DECIMAL +/- 0.1 .XX DECIMAL +/- 0.02 .XXX DECIMAL +/-0.005
 ANGLES +/- 1 DEG SURFACE FINISH 128 microinch RMS



DESCRIPTION PAVE-Seal@ Cable Harnesses
NW40-SS-150-12P-BXT20-48-48

PART NUMBER
2441

