



NOTES:

1. LEAK TEST: 80 PSI, NO BUBBLES 8 MINUTES MINIMUM, PRESSURE ON VACUUM SIDE OF HOUSING.
2. HYPOT 100 VDC 50 MEGOHMS MINIMUM 0.01 SECOND MINIMUM WIRE TO WIRE AND HOUSING.
3. ALL TESTS ARE PERFORMED AT ROOM TEMPERATURE.
4. ALL PARTS MUST PASS ALL TESTS.
5. WIRE POSITIONING IS APPROXIMATE & VARIABLE.
6. VOIDS SMALLER THAN ϕ .035 [.89] ARE ACCEPTABLE.
7. DIMENSION IS OVER-ALL-LENGTH EXCLUDING OVERFILL ON EACH SIDE OF HOUSING.
8. REF-OPERATING TEMPERATURE -20°C TO 150°C.
9. DIMENSIONS ARE INCHES [millimeters].

4	16	BRADY MARKERS	BRADY MARKERS 1-8
3	8	TT24	CABLE T/C TYPE T#24 TEFLON
2	A/R	PAVE-Seal 150	EPOXY BLACK
1	1	0346	HOUSING PT8 HP
ITEM	QTY	PART NUMBER	DESCRIPTION

PAVE technology co.

2751 Thunderhawk Court
Dayton, OH 45414-3445
U.S.A.
tel (937) 890-1100
fax (937) 8905165
www.pavetechnologyco.com

DESCRIPTION
PT8-SS-150-8P-TT24-120-40

PART NUMBER
3285

MATERIAL
NOTED

REVISION LEVEL
D

PROJECTION

ALL DIMENSIONS AND TOLERANCES APPLY TO FINISHED PART IN INCHES.
ALLOWABLE TOLERANCES UNLESS SPECIFIED OTHERWISE: NONE ± 0.5
X.X DECIMAL ± 0.1 X.XX DECIMAL ± 0.02 X.XXX DECIMAL ± 0.005
ANGLES ± 1 DEGREE SURFACE FINISH 128 microinch RMS