


NOTES:

1. APPLY VAC-GOOP TO O-RING AND WIPE OFF EXCESS WITH KIMWIPE.
2. REF- INSERTION DEPTH OF MHV CONNECTION IS .394".
3. ELECTRICAL TEST PIN TO SOCKET CONTINUITY - 1 OHM MAX
4. HYPOT TEST- CENTER PIN TO SHIELD & HOUSING 630 VDC 500 MEG OHMS 10 mS.

		2751 Thunderhawk Court Dayton, Ohio 45414-3445 U.S.A. (937)890-1100 fax (937)890-5165 www.pavetechnologyco.com
TITLE OR PAVE ORDER CODE		
MHV PLUG ASY (93ohm)		
PART NUMBER	PROJECTION	
PT-1090	