



**NOTES:**

1. LEAK TEST: 80 PSI, NO BUBBLES 2 MINUTES MIN.
2. HYPOT 630 VDC 500 MEGOHMS .01 SEC MINIMUM, WIRE TO WIRE & HOUSING.
3. ITEM 2 POSITIONS EXITING ITEM 1 PRESSURE SIDE ARE RANDOM.
4. ALL TESTS ARE DONE AT ROOM TEMPERATURE.
5. THESE SPECIFICATIONS SUPERSEDE PAVE CATALOG.

5	A/R	SUPER-O-LUBE	LUBRICANT O-RING SILICONE BASED
4	1	-012 VITON	O-RING -012 VITON 75
3	A/R	PAVESEAL 150	EPOXY BLACK
2	5x36.5	KT24	CABLE T/C TYPE K 24AWG TEFLON
1	1	1226-S	HOUSING SP8S-E
ITEM	QTY	PART NUMBER	DESCRIPTION

**PAVE technology co.**

2751 Thunderhawk Court  
Dayton, Ohio 45414-3445  
U.S.A.  
(937)890-1100  
fax (937)890-5165  
www.pavetechnologyco.com

TITLE OR PAVE ORDER CODE  
**SP8S-E-150-5P-KT24-36-36**

PART NUMBER  
**1896**

PROJECTION