


**NOTES:**

1. HE LEAK TEST @1 ATM. <math>10^{-6}</math> CC/S.
2. HYPOT 630 VDC 500 MW .01 SEC MINIMUM, WIRE TO WIRE & HOUSING
3. ALL TESTS ARE PERFORMED AT ROOM TEMPERATURE.
4. ALL SPECIFICATIONS AND TESTS SUPERSEDE PAVE CATALOG SPECIFICATIONS.

3	A/R	PAVE-Seal 150	EPOXY BLACK
2	6@50"	KT20	WIRE 20AWG TEFLON T/C TYPE K
1	1	NW-25FL	HOUSING NW25-SS
ITEM	QTY	PART#	DESCRIPTION

ALL DIMENSIONS AND TOLERANCES APPLY TO FINISHED PART IN INCHES  
 ALLOWABLE TOLERANCES UNLESS SPECIFIED OTHERWISE: NONE +/- .5  
 .X DECIMAL +/- .1 .XX DECIMAL +/- .02 .XXX DECIMAL +/- .005  
 ANGLES +/- 1 DEG SURFACE FINISH 128 microinch RMS



2751 Thunderhawk Court  
 Dayton, Ohio 45414-3445  
 U.S.A.  
 (937) 890-1100  
 fax (937) 890-5165  
 www.pavetechnologyco.com

DESCRIPTION PAVE-Seal® Cable Harnesses  
**NW25-SS-150-6P-KT20-24-24**

PART NUMBER	PROJECTION
1960	