



NOTES:

1. LEAK TEST: 750 PSI, NO BUBBLES 1 MINUTE MINIMUM.
2. HYPOT 630 VDC 500 MEGOHMS MINIMUM 0.01 SECOND MINIMUM WIRE TO WIRE AND HOUSING.
3. ALL TESTS ARE PERFORMED AT ROOM TEMPERATURE.
4. ALL PARTS MUST PASS ALL TESTS.
5. VOIDS LARGER THAN $\phi .035$ [.89] ARE NOT ACCEPTABLE.
6. REF-OPERATING TEMPERATURE -20°C TO 125°C .
7. DIMENSIONS ARE INCHES [millimeters].

3	1	2100-260	CABLE AXOPATCH 200B PVC JACKET 2/ET20+8/ET28+7/ET28/SHIELDED WITH MALE DB25 CONNECTOR & BACKSHELL
2	A/R	PAVE-Seal 150	EPOXY BLACK
1	1	1625-B	HOUSING PT16-B
ITEM	QTY	PART NUMBER	DESCRIPTION

2751 Thunderhawk Court
Dayton, OH 45414-3445
U.S.A.
tel (937) 890-1100
fax (937) 8905165
www.pavetechnologyco.com

PAVE technology co.

DESCRIPTION
PT16-B-150-1-AXOPATCH200B-70-48

PART NUMBER 3618	MATERIAL NOTED
REVISION LEVEL A	PROJECTION

ALL DIMENSIONS AND TOLERANCES APPLY TO FINISHED PART IN INCHES.
ALLOWABLE TOLERANCES UNLESS SPECIFIED OTHERWISE: NONE ± 0.5
X.X DECIMAL ± 0.1 X.XX DECIMAL ± 0.02 X.XXX DECIMAL ± 0.005
ANGLES ± 1 DEGREE SURFACE FINISH 128 microinch RMS