



**NOTES:**

1. LEAK TEST: 200 PSI NITROGEN, THREAD SIDE, 5 MINUTES MIN, NO VISIBLE BUBBLES.
2. HELIUM LEAK TEST: LESS THAN 1x10<sup>-5</sup> CC/SEC.
3. ELECTRICAL TEST: 630 VDC, 500 MEGOHMS MIN, WIRE TO WIRE AND HOUSING.
4. WIRE PAIRS TO HAVE 3 TO 4 TWISTS PER INCH.

6	A/R	PARKER SUPER-O-LUBE	VACUUM GREASE
5	A/R	PAVESEAL 150	EPOXY
4	1	3-0912 75F	O-RING, VITON, 75A
3	1	1386	SO16-SS HOUSING
2	4	28AWG TYPE EE TEFLON	WIRE, TWISTED PAIRS, COLORS: RED & BLACK
1	1	22AWG TYPE E TEFLON	WIRE, COLOR: YELLOW
ITEM	QTY	PART NUMBER	DESCRIPTION



TITLE OR PAVE ORDER CODE

**SO16-SS-150-4TP/1-TEE28/TE22**

PART NUMBER

**1384**

PROJECTION

