



NOTES:

1. LEAK TEST: 1000 PSI N2, NO BUBBLES 1 MINUTE MINIMUM.
2. HYPOT 630 VDC 500 MEGOHMS MINIMUM 0.01 SECOND MINIMUM WIRE TO WIRE.
3. ELECTRICAL TEST: 300 VAC 0.5mA MAXIMUM LEAKAGE 1 SECOND MINIMUM WIRE TO WIRE.
4. ALL TESTS ARE PERFORMED AT ROOM TEMPERATURE.
5. ALL PARTS MUST PASS ALL TESTS.
6. WIRE POSITIONING IS APPROXIMATE & VARIABLE.
7. NO VOIDS ARE ACCEPTABLE ON O-RING SEALING SURFACES, VOIDS SMALLER THAN $\phi .035$ [.89] ARE ACCEPTABLE ON ALL OTHER SURFACES.
8. DIMENSION IS OVER-ALL-LENGTH EXCLUDING OVERFILL ON ATMOSPHERE SIDE OF HOUSING.
9. REF-OPERATING TEMPERATURE -20°C TO 125°C .
10. REF-PARTS ARE DEFLASHED ON PARTING LINE $+.005$ [.13] MAXIMUM.
11. DIMENSIONS ARE INCHES [millimeters].

12	A/R	PARKER O-LUBE	LUBRICANT O-RING BARIUM-BASED
11	2	-015 VITON	O-RING -015 VITON 75
10	1	E22 VIOLET	WIRE 16E MIL-W-16878/4 .010 19/#29
9	1	E22 ORANGE	WIRE 16E MIL-W-16878/4 .010 19/#29
8	1	E22GRAY	WIRE 16E MIL-W-16878/4 .010 19/#29
7	1	E22 BROWN	WIRE 16E MIL-W-16878/4 .010 19/#29
6	1	E22 BLUE 7 STR	WIRE 16E MIL-W-16878/4 .010 19/#29
5	1	E22 BLK	WIRE 16E MIL-W-16878/4 .010 19/#29
4	1	EE10 RED	WIRE 10EE MIL-W-16878/5 .017 37/#26
3	1	EE10 GREEN	WIRE 10EE MIL-W-16878/5 .017 37/#26
2	A/R	PAVE-Seal 150	EPOXY BLACK
1	1	1374L	HOUSING SP11L-E
ITEM	QTY	PART NUMBER	DESCRIPTION

PAVE technology co.

2751 Thunderhawk Court
Dayton, OH 45414-3445
U.S.A.
tel (937) 890-1100
fax (937) 8905165
www.pavetechnologyco.com

DESCRIPTION
SP11L-E-150-2/6-TEE10/TE22-24-36

PART NUMBER
3554

MATERIAL NOTED

REVISION LEVEL
A

PROJECTION

ALL DIMENSIONS AND TOLERANCES APPLY TO FINISHED PART IN INCHES.
ALLOWABLE TOLERANCES UNLESS SPECIFIED OTHERWISE: NONE ± 0.5
X.X DECIMAL ± 0.1 X.XX DECIMAL ± 0.02 X.XXX DECIMAL ± 0.005
ANGLES ± 1 DEGREE SURFACE FINISH 128 microinch RMS