

## NOTES:

- 1. HE LEAK TEST @ 1 ATM 10x10^-8 CC/SEC OR LESS.
- 2. TEST IS PERFORMED AT ROOM TEMPERATURE.
- 3. ALL PARTS MUST PASS TEST.
- 4. VOIDS LARGER THAN  $\emptyset$ .035 [.89] ARE NOT ACCEPTABLE.
- 5. REF-OPERATING TEMPERATURE -20°C TO 125°C.
- 6. DIMENSIONS ARE INCHES [millimeters].
- 7. PAVE-SEAL CAN BE A BI-DIRECTIONAL HERMETIC SEAL FOR VACUUM AND MOST PRESSURES. FOR PRESSURES ABOVE 150 PSI (10 BAR), CHECK WITH SALES ENGINEERING.

ITEM	QTY	PART NUMBER	DESCRIPTION
1	1	1531	HOUSING VS26-SS
2	A/R	PAVE-Seal 150	EPOXY BLACK
3	1	-226 VITON	O-RING -226 VITON 75
4	1	0185	NUT VS26-SS



2751 Thunderhawk Court Dayton, OH 45414-3445 U.S.A. tel (937) 890-1100 fax (937) 8905165

VS26-SS-150-PLUG

4022

NOTED PROJECTION

ALL DIMENSIONS AND TOLERANCES APPLY TO FINISHED PART IN INCHES. ALLOWABLE TOLERANCES UNLESS SPECIFIED OTHERWISE: NONE ±0.5 X.X DECIMAL ±0.1 X.XX DECIMAL ±0.02 X.XXX DECIMAL ±0.005 ANGIES +1 DEGREE SURFACE FINISH 128 microinch RMS