



- NOTES:
1. A MINIMUM OF 4 WIRE COLORS SHALL BE USED.
 2. ALL WIRES SHALL BE TAGGED WITH WIRE MARKERS ON BOTH ENDS.
 3. WIRE ENDS MAY BY STRIPPED .25" MAX

VACUUM SIDE -
HELIUM LEAK TEST TO 1×10^{-8} CC/S

ELECTRICAL TEST -
630 VDC TO 500 MEGOHMS MIN
FOR 10 mS MIN WIRE TO WIRE
AND HOUSING.

2751 Thunderhawk Court
Dayton, Ohio 45414-3445
U.S.A.
(937)890-1100
fax (937)890-5165
www.pave technology.co

PAVE technology co.

TITLE OR PAVE ORDER CODE
PAVESEAL VS26 20 TE16

PART NUMBER
0915

PROJECTION