


3	A/R	PAVESEAL 200	EPOXY
2	1	PAVE 1479	PTN6-SS HOUSING
1	4	OMEGA RSX-24-TT	RX/RS TEFLON T/C CABLE
ITEM	QTY	PART NUMBER	DESCRIPTION

NOTES:



1. LEAK TEST, 80 PSI 36" SIDE, 10 MINUTES MIN, NO VISIBLE BUBBLES.
2. ELECTRICAL TEST: 630 VDC, 500 MEGOHMS MIN, WIRE TO WIRE & HOUSING.



2751 Thunderhawk Court
Dayton, Ohio 45414-3445
U.S.A.
(937)890-1100
fax (937)890-5165
www.pavetechnologyco.com

TITLE OR PAVE ORDER CODE

PTN6-SS-200-4P-RT24-36-204

<p style="font-size: x-small;">PART NUMBER</p> <p style="font-size: large; color: blue; font-weight: bold;">1478</p>	<p style="font-size: x-small;">PROJECTION</p> <div style="display: flex; justify-content: space-around; align-items: center;">   </div>
--	---