



NOTES:

1. HELIUM LEAK TEST@ 1 ATMOSPHERE, < 1 X 10⁻⁸ cc/s, AIR EQ.
2. HYPOT 630 VDC, 500 M , .01 SBC MIN. , ALL WIRES, BRAID & HOUSING



2751 Thunderhawk Court
Dayton, Ohio 45414-3445
U.S.A.
(937)890-1100
fax (937)890-5165
www.pave technologyco.com

TITLE OR PAVE ORDER CODE

VS32-SS-150-95TP-TE20

PART NUMBER

1496

PROJECTION

