



NOTES:

1. HIPOT TEST: 200 VDC, 100 MEGOHMS, .01 S MIN.
2. HELIUM LEAK TEST: LESS THAN 1×10^{-7} cc/s AIR EQ.

Design Library



TITLE OR PAVE ORDER CODE

KF40-SS-150-32TSP-CWNMUF-96-36

PART NUMBER

1510

PROJECTION

