



NOTES:

1. LEAK TEST @ 120 PSI FOR 3 MINUTES- NO BUBBLES.
2. ELECTRICAL HYPOT ALL WIRES & HOUSING 630VDC 500 MEGOHM .01S

TITLE OR PAVE ORDER CODE	
PT24-SS-150-20P-TT24-36-36	
PART NUMBER	PROJECTION
1683	