



ATMOSPHERE VACUUM



NOTES:

1. HE LEAK TEST @1 ATM. 10^{-6} CC/S AIR EQUIVALENT.
2. HIPOT: 630 VDC 500 MEGOHMS .01SEC W/MATING PLUGS, CENTER CONDUCTOR TO SHIELD TO HOUSING. (SHIELD IS FLOATING)
3. CONTINUITY TEST W/MATING PLUGS -1.0 OHM MAX, CENTER CONDUCTOR & SHIELD, END TO END

PAVE technology co.
 2751 Thunderhawk Court
 Dayton, Ohio 45414-3445
 U.S.A.
 (937)890-1100
 fax (937)890-5165
 www.pavetechnologyco.com

TITLE OR PAVE ORDER CODE

VS18-SS-150-BNC-BNC

PART NUMBER

1750

PROJECTION

