



**NOTES:**

1. HELIUM LEAK TEST TO  $1 \times 10^{-7}$  cc/sec @ 1 ATMOSPHERE
2. ELECTRICAL TEST, 630 VDC 500 MEGOHMS MIN.

6	1	0199	NUT VS22-SS
5	A/R	PAVE SEAL 150	EPOXY
4	A/R	SUPER-O-LUBE	LUBRICANT O-RING SILICONE BASED
3	1	-224 VITON 75A	O-RING -224 VITON 75A
2	1	0788	HOUSING VS22-SS
1	18@99"	KT20	WIRE 20AWG TEFLON T/C TYPE K
ITEM	QTY.	ITEM NUMBER	DESCRIPTION



TITLE OR PAVE ORDER CODE  
**VS22-SS-150-18P-KT20-48-48**

PART NUMBER  
**1779**

