


NOTES:


1. HE LEAK TEST @ 1 ATM. 10×10^{-7} CC/S OR LESS.
2. HYPOT 630 VDC 500 MEGAOHMS MINIMUM 0.01 SECOND MINIMUM WIRE TO WIRE AND HOUSING.
3. CONTINUITY TEST <3 Ohms PIN TO WIRE END.
4. ALL TESTS ARE PERFORMED AT ROOM TEMPERATURE.
5. ALL PARTS MUST PASS ALL TESTS.
6. WHITE INSULATED WIRE IS CONNECTED TO THE + SIDE OF THE CONNECTOR, RED TO THE - SIDE.
7. REF-OPERATING TEMPERATURE RANGE FOR:
 - a. PAVE HOUSING SEAL -20° TO 150°C.
 - b. FIT-221 SHRINK TUBING -55°C TO 135°C.
 - c. PCLM-SMP-FT FERRITE BEAD IS 120°C MAXIMUM.
 - d. FIT-105 SHRINK TUBING -20°C TO 105°C.
8. DIMENSIONS ARE INCHES [millimeters].

10	A/R	FIT-105-1	TUBING SHRINK 1" IRR. PVC
9	1	PCLM-SMP-FT	CLAMP CABLE FERRITE
8	A/R	FIT-221-3/8	TUBING SHRINK 3/8" IRR POLY
7	A/R	FIT-221-1	TUBING SHRINK 1" IRR. POLY
6	3	WIRE TIE 820/C	WIRE TIE NYLON NICHIFU
5	3	GMP-J-M	CONNECTOR MINI BLACK J TYPE
4	1	TUBULAR BRD 3/8	TUBULAR BRAID 3/8" ALPHA 2172
3	A/R	PAVE-Seal 200	EPOXY BLACK
2	3	JT20	CABLE T/C TYPE J 20AWG TEFLON
1	1	NW16	HOUSING NW16-SS
ITEM	QTY	PART NUMBER	DESCRIPTION


 2751 Thunderhawk Court
 Dayton, OH 45414-3445
 U.S.A.
 tel (937) 890-1100
 fax (937) 8905165
 www.pavetechnologyco.com

DESCRIPTION
 NW16-SS-200-3P-JT20-104-4

PART NUMBER
2622

MATERIAL NOTED
 REVISION LEVEL H
 PROJECTION 

ALL DIMENSIONS AND TOLERANCES APPLY TO FINISHED PART IN INCHES.
 ALLOWABLE TOLERANCES UNLESS SPECIFIED OTHERWISE: NONE ±0.5
 X.X DECIMAL ±0.1 X.XX DECIMAL ±0.02 X.XXX DECIMAL ±0.005
 ANGLES ±1 DEGREE SURFACE FINISH 128 microinch RMS