



NOTES:

- HE LEAK TEST VACUUM 10×10^{-7} CC/SEC OR LESS.
- ITEM 6 MUST BE REMOVED PRIOR TO INSTALLATION IN PORT.
- HYPOT 100 VDC 5 MEGOHMS MINIMUM 0.01 SECOND MINIMUM WIRE TO WIRE AND HOUSING.
- CONTINUITY TEST CONNECTOR TO CONNECTOR:
POSITIVE TERMINALS OF ITEM 5 TO ITEM 6 $< 8\Omega$ (CHROMEL)
NEGATIVE TERMINALS OF ITEM 5 TO ITEM 6 $< 3.5\Omega$ (ALUMEL)
- ALL TESTS ARE PERFORMED AT ROOM TEMPERATURE.
- ALL PARTS MUST PASS ALL TESTS.
- WIRE POSITIONING IS APPROXIMATE & VARIABLE.
- VOIDS LARGER THAN $\phi 0.035$ [.89] ARE NOT ACCEPTABLE.
- REF-OPERATING TEMPERATURE -20°C TO 155°C .
- DIMENSIONS ARE INCHES [millimeters].
- PAVE-SEAL CAN BE A BI-DIRECTIONAL HERMETIC SEAL FOR VACUUM AND MOST PRESSURES. FOR PRESSURES ABOVE 150 PSI (10 BAR), CHECK WITH SALES ENGINEERING.

ALL DIMENSIONS AND TOLERANCES APPLY TO FINISHED PART IN INCHES.
ALLOWABLE TOLERANCES UNLESS SPECIFIED OTHERWISE: NONE ± 0.5
X.X DECIMAL ± 0.1 X.XX DECIMAL ± 0.02 X.XXX DECIMAL ± 0.005
ANGLES ± 1 DEGREE SURFACE FINISH 128 microinch RMS

2751 Thunderhawk Court
Dayton, OH 45414-3445
U.S.A.
tel (937) 890-1100
fax (937) 8905165
www.pavetechnologyco.com

PAVE technology co.

DESCRIPTION
VS15L-SS-200-6P-KR20-96T-96T

PART NUMBER
2667

MATERIAL NOTED

REVISION LEVEL -

PROJECTION