



NOTES:

1. LEAK TEST: 80 PSI, NO BUBBLES 3 MINUTES MINIMUM.
2. HYPOT 100 VDC 10 MEGAOHMS MINIMUM 0.01 SECOND MINIMUM WIRE TO WIRE AND HOUSING.
3. ALL TESTS ARE PERFORMED AT ROOM TEMPERATURE.
4. ALL PARTS MUST PASS ALL TESTS.
5. REF- OPERATING TEMPERATURE: -20 TO 200 °C.

3	12	KG20	CABLE T/C TYPE K 20AWG GLASS
2	A/R	PAVE-Seal 200	EPOXY BLACK
1	1	1132	HOUSING PTN16-SS
<b>ITEM</b>	<b>QTY</b>	<b>PART NUMBER</b>	<b>DESCRIPTION</b>

**PAVE technology co.**  
 2751 Thunderhawk Court  
 Dayton, OH 45414-3445  
 U.S.A.  
 tel (937) 890-1100  
 fax (937) 8905165  
 www.pavetechnologyco.com

DESCRIPTION  
 PTN16-SS-200-12P-KG20-36-36

PART NUMBER  
**3115**

REVISION LEVEL -  
 PROJECTION

ALL DIMENSIONS AND TOLERANCES APPLY TO FINISHED PART IN INCHES.  
 ALLOWABLE TOLERANCES UNLESS SPECIFIED OTHERWISE: NONE ±0.5  
 X.X DECIMAL ±0.1 X.XX DECIMAL ±0.02 X.XXX DECIMAL ±0.005  
 ANGLES ±1 DEGREE SURFACE FINISH 128 microinch RMS