

NOTES:

- 1. LEAK TEST: 5,000 PSI, NO BUBBLES 2 MINUTES MINIMUM. ONLY FIBERS IN PT-8 HOUSING TESTED THEN ASSEMBLED IN TUBE.
- 2. TUBE AND SWAGELOK FITTING RATED TO 5000 PSI.
- 3. OPTICAL CONTINUITY TEST, PART MUST CONTAIN 10 MINIMUM FIBERS WITH OPTICAL CONTINUITY AFTER ASSEMBLED IN TUBE.
- 4. ALL TESTS ARE PERFORMED AT ROOM TEMPERATURE.
- 5. ALL PARTS MUST PASS ALL TESTS.
- 6. FIBER POSITIONING IS APPROXIMATE & VARIABLE.
- 7. VOIDS SMALLER THAN \emptyset 0.035 ARE ACCEPTABLE ON ALL OTHER SURFACES.
- 8. REF-OPERATING TEMPERATURE -20 °C TO 200 °C.
- 9. DIMENSIONS ARE INCHES [MILLIMETERS].

ITEM	QTY	PART NUMBER	DESCRIPTION	3214	PROJECTION O	
1	1	3236	TUBING .25" OD 1M LONG 5000 PSI SS	3214	LEVEL -	
2	1	SS-400-7-4	SWAGELOK 1/4 PIPE TO 1/4 FEMALE NPT	PART NUMBER	REVISION	
3	1	SS-400-7-8	SWAGELOK 1/4 PIPE TO 1/2 FEMALE NPT	SO12L-SS-200-10-SMFO-2-2m		
4	1	0346	HOUSING PT-8 SS			
5	12	LXSIX FIBER	FIBER SINGLE MODE 9/125/155 POLYIMIDE	0	fax (937) 8905165 www.pavetechnologyco.com	
6	12	LXSIX TUBING	TUBING .012ID .030OD TEFLON	PAVE ledinology of	U.S.A. tel (937) 890-1100	
7	A/R	PAVE-Seal 200	EPOXY BLACK	M	Dayton, OH 45414-3445	

ALL DIMENSI	ONS ANL) TOLERANC	ES APPLY I) FINISHED	PARI IN INC	JHES.
ALLOWAB LE	TOLERAN(DES UNLESS	SPECIFIED	OTHERWISE	E: NONE ±0.	5
X.X DECIMAL	_ ±0.1	X.XX DECIN	MAL ±0.02	X.XXX DEC	IMAL ±0.005	
ANGLES ±1	DEGREE	SURFACE	FINISH 128	microinch	RMS	