



NOTES:

1. LEAK TEST: 1500 PSI, NO BUBBLES 60 SECONDS MINIMUM.
2. HYPOT 630 VDC 500 MEGAOHMS MINIMUM 0.01 SECOND MINIMUM WIRE TO WIRE.
3. ALL TESTS ARE PERFORMED AT ROOM TEMPERATURE.
4. ALL PARTS MUST PASS ALL TESTS.
5. DIMENSION IS OVER-ALL-LENGTH EXCLUDING OVERFILL ON ATMOSPHERE SIDE OF HOUSING.
6. NO VOIDS ARE ACCEPTABLE ON O-RING SEALING SURFACES, VOIDS SMALLER THAN $\phi .035$ [.89] ARE ACCEPTABLE ON ALL OTHER SURFACES.
7. REF-PARTS ARE DEFLASHED ON PARTING LINE $+.005$ [.13] MAXIMUM.
8. WIRE POSITIONING IS APPROXIMATE & VARIABLE.
9. DIMENSIONS ARE INCHES [millimeters].

8	A/R	PARKER O-LUBE	LUBRICANT BARIUM-BASED
7	2	-015 VITON	O-RING -015 VITON 75
6	2	PC22 UL1015 WHT	WIRE 22PVC UL1015 .032 7/#30
5	2	PC22 UL1015 BLK	WIRE 22PVC UL1015 .032 7/#30
4	2	TT-K-24S	CABLE T/C 24 TEFLON STRANDED
3	1	BELDEN 9L28009	CABLE PVC #28 7 STRAND .050 PITCH
2	1	PAVE-Seal 150	EPOXY BLACK
1	1	1374M	HOUSING SP11M-E
ITEM	QTY	PART NUMBER	DESCRIPTION

PAVE technology co.

2751 Thunderhawk Court
Dayton, OH 45414-3445
U.S.A.
tel (937) 890-1100
fax (937) 8905165
www.pavetechnologyco.com

DESCRIPTION SP11M-E-150-6/2P/4-RPC28/K24S/PC22-18-18	
PART NUMBER 3325	MATERIAL NOTED
REVISION LEVEL -	PROJECTION

ALL DIMENSIONS AND TOLERANCES APPLY TO FINISHED PART IN INCHES.
ALLOWABLE TOLERANCES UNLESS SPECIFIED OTHERWISE: NONE ± 0.5
X.X DECIMAL ± 0.1 X.XX DECIMAL ± 0.02 X.XXX DECIMAL ± 0.005
ANGLES ± 1 DEGREE SURFACE FINISH 128 microinch RMS