



NOTES:

1. HE LEAK TEST @ 1ATM 1X10⁻⁷ CC/SEC OR LESS. LEAK TEST IS PERFORMED USING A G-275 VITON GASKET, A GREEN PAINT DOT INDICATES THE PART HAS PASSED THE TEST.
2. HYPOT 630 VDC 500 MEGOHMS MINIMUM 0.01 SECOND MINIMUM WIRE TO WIRE AND HOUSING.
3. ALL TESTS ARE PERFORMED AT ROOM TEMPERATURE.
4. ALL PARTS MUST PASS ALL TESTS.
5. WIRE POSITIONING IS APPROXIMATE & VARIABLE.
6. COSMETIC SURFACE VOIDS SMALLER THAN Ø0.035 [0.89] ARE ACCEPTABLE.
7. DIMENSIONS ARE INCHES [millimeters].
8. REF-OPERATING TEMPERATURE -20°C TO 125°C.
9. PAVE-SEAL CAN BE A BI-DIRECTIONAL HERMETIC SEAL FOR VACUUM AND MOST PRESSURES. FOR PRESSURES ABOVE 150 PSI (10 BAR), CHECK WITH SALES ENGINEERING.

4900-1	TT24 SLE	CABLE T/C TYPE T#24 TEF SPECIAL LIMITS OF ERROR
4900	TT24	CABLE T/C TYPE T 24AWG TEFLON
PAVE #	ITEM 3	DESCRIPTION

4	1	GASKET G-275 C	GASKET COPPER
3	16	SEE TABLE	SEE TABLE
2	A/R	PAVE-Seal 150	EPOXY BLACK
1	1	1473	HOUSING CF275-SS
ITEM	QTY	PART NUMBER	DESCRIPTION

PAVE technology co.
 2751 Thunderhawk Court
 Dayton, OH 45414-3445
 U.S.A.
 tel (937) 890-1100
 fax (937) 8905165
 www.pavetechnologyco.com

DESCRIPTION
 CF275-SS-150-16P-TT24-96-48

PART NUMBER
4900*

MATERIAL NOTED

REVISION LEVEL
 A

PROJECTION

ALL DIMENSIONS AND TOLERANCES APPLY TO FINISHED PART IN INCHES.
 ALLOWABLE TOLERANCES UNLESS SPECIFIED OTHERWISE: NONE ±0.5
 X.X DECIMAL ±0.1 X.XX DECIMAL ±0.02 X.XXX DECIMAL ±0.005
 ANGLES ±1 DEGREE SURFACE FINISH 128 microinch RMS