



NOTES:

1. HE LEAK TEST @ 1 ATM. 1×10^{-6} CC/SEC OR LESS. A GREEN PAINT DOT INDICATES THE PART HAS PASSED THE TEST.
2. HYPOT 630 VDC 500 MEGOHMS MINIMUM 0.01 SECOND MINIMUM, $<0.5 \Omega$ CONTINUITY, CONDUCTOR TO SHIELD, AND HOUSING WITH MATING CONNECTORS.
3. ALL TESTS ARE PERFORMED AT ROOM TEMPERATURE.
4. ALL PARTS MUST PASS ALL TESTS.
5. CONNECTOR POSITIONING IS APPROXIMATE & VARIABLE.
6. COSMETIC SURFACE VOIDS NOT ON O-RING SEALING SURFACES ARE ACCEPTABLE BASED ON THE SEAL'S DIAMETER:
 $\leq .5 [12.7]$ SEAL DIAMETER: $\phi .035 [.89]$ MAX ALLOWED VOID SIZE
 $> .5 [12.7]$ SEAL DIAMETER: $\phi .060 [1.5]$ MAX ALLOWED VOID SIZE
7. REF-OPERATING TEMPERATURE -20°C TO 125°C .
8. DIMENSIONS ARE INCHES [millimeters].

ALL DIMENSIONS AND TOLERANCES APPLY TO FINISHED PART IN INCHES.
 ALLOWABLE TOLERANCES UNLESS SPECIFIED OTHERWISE: NONE ± 0.5
 X.X DECIMAL ± 0.1 X.XX DECIMAL ± 0.02 X.XXX DECIMAL ± 0.005
 ANGLES ± 1 DEGREE SURFACE FINISH 128 microinch RMS

3	8	5413779-3	RCPT BNC
2	A/R	PAVE-Seal 150	EPOXY BLACK
1	1	3583	HOUSING NW40-SS
ITEM	QTY	PART NUMBER	DESCRIPTION

PAVE technology co.

2751 Thunderhawk Court
 Dayton, OH 45414-3445
 U.S.A.
 tel (937) 890-1100
 fax (937) 8905165
 www.pavetechnologyco.com

DESCRIPTION
 NW40-SS-150-4-BNC-BNC

PART NUMBER
5149

MATERIAL NOTED

REVISION LEVEL -

PROJECTION