


NOTES:

1. HE LEAK TEST @ 1 ATM. 1X10⁻⁷ CC/SEC OR LESS. A GREEN PAINT DOT ON THE ATMOSPHERE SIDE INDICATES THE PART HAS PASSED THE TEST.
2. OPTICAL CONTINUITY CHECK, 12 MINIMUM GOOD FIBERS IN EACH BUNDLE OF 13 (1 SPARE PER BUNDLE), 48 TOTAL.
3. ALCOHOL OR FOMBLIN YVAC 3 MAY BE USED TO LUBRICATE O-RINGS FOR VACUUM HELIUM LEAK TESTING.
4. ALL TESTS ARE PERFORMED AT ROOM TEMPERATURE.
5. ALL PARTS MUST PASS ALL TESTS.
6. FIBER POSITIONING IS APPROXIMATE & VARIABLE.
7. FIBERS ARE NOT REQUIRED TO BEND SHARPLY AT EPOXY SEAL SURFACE.
8. COSMETIC SURFACE VOIDS NOT ON O-RING SEALING SURFACES ARE ACCEPTABLE BASED ON THE SEAL'S DIAMETER:
 <= .5 [12.7] SEAL DIAMETER: ϕ .035 [.89] MAX ALLOWED VOID SIZE
 > .5 [12.7] SEAL DIAMETER: ϕ .060 [1.5] MAX ALLOWED VOID SIZE
9. REF-OPERATING TEMPERATURE RANGE -20°C TO 65°C.
10. DIMENSIONS ARE INCHES [millimeters].
11. PAVE-SEAL CAN BE A BI-DIRECTIONAL HERMETIC SEAL FOR VACUUM AND MOST PRESSURES. FOR PRESSURES ABOVE 150 PSI (10 BAR), CHECK WITH SALES ENGINEERING.
12. 5272-1 ONLY IS TERMINATED, TO THE LENGTH SHOWN, ON BOTH SIDES WITH ITEM 8. PINNING IS 1 TO 1, & MALE TO MALE. AN INSERTION LOSS TEST IS PERFORMED @ 850nm, .35dB OR BETTER ALL FIBERS AND BOTH DIRECTIONS WITH MATING CONNECTORS.

SHIPPED WITH PROTECTIVE CAPS INSTALLED, REMOVED FOR CLARITY BOTH SIDES

SPARE FIBERS FROM EACH BUNDLE ARE NOT TERMINATED

ITEM	QTY	PART NUMBER	DESCRIPTION
9	SEE NOTE 12 8	USC23834	CONNECTOR MTP 12 POS MM FERRULE MALE DEMAGNETIZED USCONEC OR EQUAL
8	4	ZUSA-HT-3406	TUBING 1/8" ID X 1/4" OD POLYETHYLENE NATURAL
7	1	1014	NUT VS18-SS
6	1	-219 EPDM	O-RING -219 EPDM 70
5	4	PTN0.25BK	SLEEVING 1/4" BRAIDED EXPANDABLE FLEXO-PET BLACK
4	4	PE021-N	TUBING .062 ID X .125 OD POLYETHYLENE NATURAL
3	**SEE NOTE 2 52*	K5C000040	FIBER OPTIC MULTIMODE 50/125um DRAKA SUPER RADHARD
2	A/R	PAVE-Seal 150	EPOXY BLACK
1	1	0660	HOUSING VS18-SS



ALL DIMENSIONS AND TOLERANCES APPLY TO FINISHED PART IN INCHES.
 ALLOWABLE TOLERANCES UNLESS SPECIFIED OTHERWISE: NONE \pm 0.5
 X.X DECIMAL \pm 0.1 X.XX DECIMAL \pm 0.02 X.XXX DECIMAL \pm 0.005
 ANGLES \pm 1 DEGREE SURFACE FINISH 128 microinch RMS

PAVE technology co.
 2751 Thunderhawk Court
 Dayton, OH 45414-3445
 U.S.A.
 tel (937) 890-1100
 fax (937) 8905165
 www.pavetechnologyco.com

DESCRIPTION
 VS18-SS-150-48-MMFO-118-118

PART NUMBER
5272

MATERIAL NOTED

REVISION LEVEL
 C

PROJECTION