



**NOTES:**

1. HE LEAK TEST @ 1 ATM.  $1 \times 10^{-7}$  CC/SEC OR LESS. A GREEN PAINT DOT ON THE ATMOSPHERE SIDE INDICATES THE PART HAS PASSED THE TEST.
2. HYPOT 1000 VDC 500 MEGOHMS MINIMUM 0.01 SECOND MINIMUM PIN TO PIN AND HOUSING WITH MATING CONNECTORS.
3. CONTINUITY TEST <0.5 OHM PIN TO SOCKET AND RECEPTACLE SHELL TO RECEPTACLE SHELL WITH MATING CONNECTORS.
4. ALCOHOL OR FOMBLIN YVAC 3 MAY BE USED TO LUBRICATE O-RINGS FOR VACUUM HELIUM LEAK TESTING.
5. ALL TESTS ARE PERFORMED AT ROOM TEMPERATURE.
6. ALL PARTS MUST PASS ALL TESTS.
7. COSMETIC SURFACE VOIDS NOT ON O-RING SEALING SURFACES ARE ACCEPTABLE BASED ON THE SEAL'S DIAMETER:  
 $\leq .5$  [12.7] SEAL DIAMETER:  $\phi .035$  [.89] MAX ALLOWED VOID SIZE  
 $> .5$  [12.7] SEAL DIAMETER:  $\phi .060$  [1.5] MAX ALLOWED VOID SIZE
8. REF-OPERATING TEMPERATURE RANGE  $-20^{\circ}\text{C}$  TO  $125^{\circ}\text{C}$ .
9. DIMENSIONS ARE INCHES [millimeters].
10. PAVE-SEAL CAN BE A BI-DIRECTIONAL HERMETIC SEAL FOR VACUUM AND MOST PRESSURES. FOR PRESSURES ABOVE 150 PSI (10 BAR), CHECK WITH SALES ENGINEERING.

6	1	0199	NUT VS22-SS
5	1	-224 VITON	O-RING -224 VITON 75
4	1	20GE26SN	RCPT 26#20 SOCKET NICKEL PLATED ALUMINUM SHELL GOLD PLATED COPPER ALLOY CONTACTS SPACE GRADE
3	1	20GE26PN	RCPT 26#20 PIN NICKEL PLATED ALUMINUM SHELL GOLD PLATED COPPER ALLOY CONTACTS SPACE GRADE
2	A/R	PAVE-Seal 150	EPOXY BLACK
1	1	0189/304SS	HOUSING VS22-SS
ITEM	QTY	PART NUMBER	DESCRIPTION


 2751 Thunderhawk Court  
 Dayton, OH 45414-3445  
 U.S.A.  
 tel (937) 890-1100  
 fax (937) 8905165  
[www.pavetechnologyco.com](http://www.pavetechnologyco.com)

DESCRIPTION  
 VS22-SS-150-26-20-99920PS (NP)

PART NUMBER	MATERIAL	NOTED
<b>5715</b>	REVISION LEVEL	A
	PROJECTION	

ALL DIMENSIONS AND TOLERANCES APPLY TO FINISHED PART IN INCHES.  
 ALLOWABLE TOLERANCES UNLESS SPECIFIED OTHERWISE: NONE  $\pm 0.5$   
 X.X DECIMAL  $\pm 0.1$  X.XX DECIMAL  $\pm 0.02$  X.XXX DECIMAL  $\pm 0.005$   
 ANGLES  $\pm 1$  DEGREE. SURFACE FINISH 128 microinch RMS